

Date Received: _____

Certificate of Analysis

Rev 0

Page 1 of 1

Catalog No.:	Lot No.:	Storage:	Solvent:	Exp. Date:	Description:
120076-01	414391	≤ -10 °C	P/T Methanol	16-Jan-2024	8260 Internal Standard Solution, 2,000 mg/L, 1 mL

Compound	CAS No.	Purity (%)	Neat Material Lot No.	Concentration
chlorobenzene-d ₅	3114-55-4	99.41	128.9.3P	1999 ± 6.93 mg/L
1,4-dichlorobenzene-d4	3855-82-1	100	3.12.6.1P	2004 ± 7.23 mg/L
1,4-difluorobenzene	540-36-3	99.5	129.158.1P	2002 ± 21.19 mg/L
pentafluorobenzene	363-72-4	99.9	131.7.1P	2001 ± 7.22 mg/L



Certified By: _____

Kayla Coleman

Manufacture Date 17-Jan-2019

Follow all storage requirements, keep tightly closed when not in use, and use good laboratory practices when handling.
This Reference Material was manufactured, produced, and/or certified under a quality management system that is accredited to ISO 17034 and ISO/IEC 17025.

All weights are traceable through N. I. S. T. Test No. 822/264157-00. Concentration (correct for purity) and uncertainty (95% confidence) values listed are determined gravimetrically.

The stated uncertainty is the expanded uncertainty with a coverage factor of two to give a 95% confidence level.